


Gilet de buste échancré

POUR TOUS FAUTEUILS

Descriptif

Il procure un confort renforcé du buste. Son échancrure évite la possibilité de strangulation.

- Décliné en 3 tailles S, M, L correspondantes aux tailles de vêtements.
- Périmètre maximum utile : 190 cm
- Boucle : 40 mm



CMB2HP-PF

Ce dispositif médical doit être manipulé uniquement par du personnel formé (par un professionnel de santé) et qualifié à son utilisation (destination d'usage, installation et environnement d'utilisation, …). En cas de mauvaise installation du dispositif médical, il y a un risque d'accident pour le patient.

2424 70 ®

V08.09/26

Matériaux textiles conformes à la législation REACH sur les substances indésirables

NAUSICAA Médical
12 Allée du Piot - ZA Pôle Actif
30660 Gallargues le Montueux
FRANCE

CE

Préconisation

L'usage du dispositif nécessite obligatoirement la présence et la vigilance d'une tierce personne. Ce dispositif médical doit être mis en œuvre ou faire l'objet d'une formation par un professionnel de santé.

Il procure un confort renforcé du buste. Son échancrure évite la possibilité de strangulation.

Composition

- Jersey polyester enduit de polyuréthane doublé de mousse polyuréthane
- Biais 100 % polyester
- Sangles 100 % polyester
- Boucles nylon

Mise en place

Ce dispositif médical est destiné au maintien & au positionnement du patient. Ce n'est pas un système de contention.
Avant chaque utilisation, il est important de vérifier : la présence de l'étiquette, les piqûres des sangles : qu'elles ne s'effilochent pas ou ne se défont pas, que le dispositif ne présente pas de déchirures, d'accrocs, de trous ni de signes d'effilochage (tissus et sangles d'accroche), que les boucles ne sont pas fissurées ou ne présentent pas de faiblesses d'attaches.

- Enfiler le maintien de buste par la tête. Faire passer la sangle de la ceinture au travers du passant, situé à l'extrémité des bretelles de la veste, en ceinturant le dossier du fauteuil. Ajuster les longueurs des sangles en fonction des besoins.





Recommandations importantes

- Afin d'assurer un maximum d'efficacité à l'utilisation des produits de cette gamme, il est indispensable :
  - de choisir la taille adéquate en fonction du patient
  - d'ajuster au mieux le produit sur le patient
- Ces produits ne doivent pas être mis au contact direct d'une peau lésée.

Gestion de la fin de vie du produit



Bust support belt

FOR ALL CHAIRS

Description

It provides reinforced comfort of the bust. Its cut-out prevents the possibility of strangulation.

- Declined in 3 sizes S, M, L corresponding to the sizes of clothing.
- Maximum useful area: 190 cm
- Buckle: 40 mm

This medical device must be handled only by employees trained (by a healthcare professional) and qualified for its use (intended purpose, installation and environment of use, etc.). If the medical device is incorrectly installed, there is a risk of accident for the patient.

CMB2HP-PF

V08.09/26

Matériaux textiles conformes à la législation REACH sur les substances indésirables

NAUSICAA Médical
12 Allée du Piot - ZA Pôle Actif
30660 Gallargues le Montueux
FRANCE

Recommendation

The use of the device requires the presence and vigilance of a third person. This medical device must be implemented or trained by a health professional.

It provides reinforced comfort of the bust. Its cut-out prevents the possibility of strangulation.

Composition

- Combination of polyurethan foam and on both faces
- Jersey/PU coating
- 100% polyester bias
- 100% polyester strap
- Nylon buckle

Set up

This medical device is intended for patient positioning.This is not a restraint system.
Before each use, it is important to check the presence of the label, the stitches of the straps, that they do not fray or break, that the device does not show any tears, snags, holes or signs of fraying (fabric and straps), that the buckles are not cracked or have no weak fasteners.

- Put on the breast support through the head.
- Pass the belt strap through the loop which is situated at the end of the support braces and grasp the backrest of the chair.
- Adjust length of straps depending on the needs.





Important recommandations

- In order to get a maximum efficiency of these sort of product, it is necessary :
- to choose the appropriate size for the patient
- to get the best possible adjustment to the patient
- These products must not be in direct contact with a wounded skin.

End-of-life management