



www.nausicaa-medical.com

This medical device is intended for patient holding and positioning. It is not a restraint system. Before each use, it is important to check: the presence of the label, the stitches of the straps: that they do not fray or break, that the device does not show tears, snags, holes or signs of fraying (fabric and straps), that the buckles are not cracked or have no weak fastening.

YZA BELT

FOR ALL CHAIRS



A one-size-fits-all abdominal belt, plus two modules (one pelvic and one bust), give you four configurations (abdominal support, bust support, pelvic support and full support).

This means you can easily switch from one to the other, adapting to all patients without size constraints.

- Maximum useful perimeter: 200 cm



YZA2CAHP-PF



The use of the device requires the presence and vigilance of a third person. When used by an agitated or disoriented person, monitoring becomes imperative. This medical device must be implemented or trained by a health professional.



v 04.06/23

Important Recommendations :

In order to get a maximum efficiency of these sort of product, it is necessary :

- to choose the appropriate size for the patient
- to get the best possible adjustment to the patient

These products must not be in direct contact with a wounded skin.



YZA BELT

HOSPITAL RANGE

Lined polyurethane coated polyester jersey
polyurethane foam
Bias binding 91% polyamide - 9% elastane
100% polyester webbing
PVC shoulder pad
Nylon buckle

INSTRUCTION FOR USE

The YZA Belt is declined under 4 modular versions :

1. the Abdominal YZA Belt.

- a - Position the belt at waist level.
- b - Pass the straps behind the back of the chair.
- c - Close the nylon buckle, then adjust the length of the straps, taking care not to compress the patient's abdomen.

2. the Bust YZA Belt is composed of the Abdominal YZA Belt to which is added a bust modular element.

- a - Position the pelvic support on the seat of the chair, and attach the 25 mm strap to the back of the chair with minimum play.
- b - Sit the patient on the chair and on the strap.
- c - Pass the straps behind the back of the chair.
- d - Close the nylon buckle, then adjust the strap length, taking care not to compress the patient's abdomen.
- e - Close the 50 mm pelvic support buckle, adjusting its length.

3. the Pelvic YZA Belt is composed of the Abdominal YZA Belt to which is added a pelvic modular element.

- a - Position the belt at waist level.
- b - Pass the straps behind the back of the chair.
- c - Close the nylon buckle, then adjust the length of the straps, taking care not to compress the patient's abdomen.
- d - Fit the shoulder harness, close the 50 mm buckle, position the shoulder straps on the patient's shoulders, with the buckles behind the back of the chair. Pass a shoulder strap under the abdominal strap, connect the buckles and adjust its length.

4. the Integral YZA Belt is composed of the Abdominal YZA Belt to which is added a bust modular element and a pelvic modular element.

- a - To fit the YZA Integral belt, proceed first as for the YZA Pelvic belt (2), then as described in paragraphs d and e for the YZA Chest belt (3).