



www.nausicaa-medical.com

This medical device is intended for patient holding and positioning. It is not a restraint system.
Before each use, it is important to check: the presence of the label, the stitches of the straps: that they do not fray or break, that the device does not show tears, snags, holes or signs of fraying (fabric and straps), that the buckles are not cracked or have no weak fastening.

PELVIC COMFORT BELT

FOR ALL CHAIRS




TO PREVENT FORWARD
SLIDING FOR ALL CHAIRS
EXCEPT SHELLS

Périmètre maximum utile : ± 230 cm
Boucle : 50 mm

Its internal STERITHANE protection makes it waterproof while being breathable.

- Available in 3 sizes S, M, L
- Maximum useful area: 230 cm
- Waist (in cm)



The use of the device requires the presence and vigilance of a third person. When used by an agitated or disoriented person, monitoring becomes imperative.
This medical device must be implemented or trained by a health professional.



CPCF3HP-PF

v 05.02/23

Important Recommendations :

- In order to get a maximum efficiency of these sort of product, it is necessary :
- to choose the appropriate size for the patient
- to get the best possible adjustment to the patient
- These products must not be in direct contact with a wounded skin.



PELVIC COMFORT BELT

HOSPITAL RANGE

Combination of polyurethan foam and
on both faces jersey/PU coating.

100% polyester bias
100% polyester strap
nylon buckle

- Place the part « pelvic holder » on the seat of the chair, the narrowest part to the backrest.
- Pass the strap of the narrowest part around the backrest of the chair, firmly adjust it and close the buckle.
- Sit the patient on the chair and raise the belt part between his legs and adjust the belt around the patient's abdomen.
- Adjust the pelvic part to avoid creases.
- Pass the strap which has not been used yet behind the backrest, close the buckle, adjust the length of the strap while being careful not to compress the abdomen of the patient.